

MEDICAL CANNULATED STAINLESS STEELS FOR IMPLANTS

| Forécreu | Specifications | | Designations | | Trademarks | |
|--------------|----------------|----------|-----------------|------------------|-------------------------------------|--|
| 316L implant | ASTM | F-138-G2 | - | - | A&D Sandvik Cartech Bohler | M25W Bioline 316LVM Biodur 316LS A220SC VMR |
| | ISO | 5832-1-D | ISO | - | | |
| | DIN | 17443 | DIN | X2CrNiMo18-14-3 | | |
| | - | - | Werkstoff N° | 1.4441 | | |
| | BS | 7252-1-D | BS | - | | |
| | | AFNOR | X2CrNiMo18-14-3 | | | |
| CN21 | ASTM | F-1586 | - | - | A&D Smp Bohler Sandvik | M30NW REX734 P504 High N |
| | ISO | 5832-9 | ISO | - | | |
| | - | - | DIN | X4CrNiMnMo21-9-4 | | |
| | BS | 7252-9 | BS | - | | |
| | - | - | AFNOR | X4CrNiMnMo21-9-4 | | |
| NoNic | ASTM | F-2229 | AISI | - | A&D Cartech | No Nic M2 Biodur 108 |
| | - | - | DIN | X8MnCrMoN23-21-1 | | |
| | - | - | Werkstoff N° | - | | |
| | - | - | AFNOR | - | | |

Slight variations between equivalent Standards are possible.